

IEC SYSTEM FOR CONFORMITY TESTING AND
CERTIFICATION OF ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

SYSTEME CEI D'ESSAIS DE CONFORMITE ET DE CERTIFICATION
DES EQUIPEMENTS ELECTRIQUES (IECEE)
METHODE OC

CB TEST CERTIFICATE

CERTIFICAT D'ESSAI OC

Product
Produit

Name and address of the applicant
Nom et adresse du demandeur

Name and address of the manufacturer
Nom et adresse du fabricant

Name and address of the factory
Nom et adresse de l'usine

Rating and principal characteristics
Valeurs nominales et caractéristiques principales

Trademark (if any)
Marque de fabrique (si elle existe)

Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais constructeur

Model / Type Ref.
Ref. de type

Additional information (if necessary)
Information complémentaire (si nécessaire)

A sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à la

as shown in the Test Report Ref. No.
which forms part of this Certificate
comme indiqué dans le Rapport d'essais numéro
de référence qui constitue partie de ce Certificat

Component - Power Supply

SL POWER ELECTRONICS CORP
6050 KING DRIVE BLDG. A
VENTURA CA 93003, USA

SL POWER ELECTRONICS CORP
6050 KING DRIVE BLDG. A
VENTURA CA 93003, USA

SL POWER ELECTRONICS XIANGHE
ANPING ECONOMIC & TECH DEVELOPING ZONE
XIANGHE HEBEI 065402, CHINA

See Information on page 2



SMT

MINT1065XY75CZ, where X is A, B, or C where Y is any number 12
through 48, where Z is any number 01 through 99, designates additional
configurations indicating non-safety related options.

This CB Test Report comprises 2 enclosures. The CB Test Certificate and Report
were amended on February 23, 2011 to add new model series, update models,
ratings and the table of critical components.

**IEC 60601-1 (1988) Second Edition,
with Amendment No. 1 (1991) and No. 2 (1995)** with the exception of:
Clause 36, Electromagnetic Compatibility, Clause 48, Biocompatibility and
Clause 52.1, Programmable Electronic Systems. Inclusive of CENELEC Common
Modifications. See Test Report for National Differences.

E116994-A44-CB-1

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme National de Certification



**Underwriters
Laboratories**

Underwriters Laboratories Inc. / GMA Certification Department, US
333 Pfingsten Road, Northbrook, IL 60062-2096
United States of America
TEL INT* +1 847 664 3008, FAX INT* +1 847 313 3008
email: jolanta.m.wroblewska@us.ul.com

Date: Issued: 2008 June 25
Amended: 2010 April 8 (Am. 1)
Amended: 2011 February 23 (Am. 2)

Signature:

Jolanta M. Wroblewska

IEC SYSTEM FOR CONFORMITY TESTING AND
CERTIFICATION OF ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

SYSTEME CEI D'ESSAIS DE CONFORMITE ET DE CERTIFICATION
DES EQUIPEMENTS ELECTRIQUES (IECEE)
METHODE OC

CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

Input: 100-240 V~, 50-60 Hz, 1.3A

Output:

Models MINT1065A Series, and MINT1065B Series - 12 Vdc, 5.25 A to 48 Vdc, 1.35 A, maximum 65 Watts.

Output:

Model MINT1065C Series - 12 Vdc, 4 A to 48 Vdc, 1 A, maximum 48 Watts.

Additional information (if necessary)
Information complémentaire (si nécessaire)

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme National de Certification




**Underwriters
Laboratories**

Underwriters Laboratories Inc. / GMA Certification Department, US
333 Pfingsten Road, Northbrook, IL 60062-2096
United States of America
TEL INT* +1 847 664 3008 FAX INT* +1 847 313 3008
email: jolanta.m.wroblewska@us.ul.com

Date: Issued: 2008 June 25
Amended: 2010 April 8 (Am. 1)
Amended: 2011 February 23 (Am. 2)

Signature:



Jolanta M. Wroblewska