

DEMKO CERTIFICATE

Certificate No. D-06890-M1
Page 1/3
Date of Issue 2019-05-10

Certificate Holder SL POWER ELECTRONICS CORP
6050 KING DRIVE, BLDG A
VENTURA, CA 93003 UNITED STATES

Manufacturer SL POWER ELECTRONICS CORP
6050 KING DRIVE, BLDG A
VENTURA, CA 93003 UNITED STATES

Production site INDUSTRIAS S L S A DE C V
CIRCUITO SIGLO XXI 2055 COL PARQUE INDUSTRIAL EX-XXI,
21254 MEXICALI BC
MEXICO

See Page 2

Certified Product Model Medical Power Supply
ME240A24YYF02, ME240B2400F02, SE240AXXYZZWW
See Page 2

Trademark



Rated Voltage / Frequency

See Page 2

Rated Current / Power

See Page 2

Insulation Class

I or II

Degree of protection (IP)

IP22

Tested acc. to

EN 60601-1-6:2010/A1:2015, EN 60601-1-6:2010, EN 60601-1:2006/A12:2014, EN 60601-1:2006/A1:2013, EN 60601-1:2006

Test Report No.

E116994-D1021-1/A1/C0-CB issued on 2019-04-29

Additional

-

Expire date

2028-11-07

Certification Manager
Jan-Erik Storgaard

This is to certify that representative sample(s) of the Product described herein ("Certified Product") have been investigated and found in compliance with the Standard(s) indicated on this Certificate, in accordance with the D Mark Requirements. The Designated Certificate holder is entitled to use the d or for cables <DEMKO> for the Certified Product manufactured at the production site(s) identified above, in accordance with the D Mark Service Agreement, including without limitation the D Mark Testing and Certification Services Service Terms. Only those Products bearing the D Mark should be considered as being covered by UL's D Mark Service. This Certificate shall remain valid through the expiration date, unless terminated earlier in accordance with the Service Agreement including without limitation if the Standard identified on this Certificate is amended or withdrawn prior the expiration date.

Certification Body

UL International Demko A/S, Borupvang 5A, DK-2750
Ballerup, Denmark, Tel. +45 44 85 65 65,
info.dk@ul.com
www.ul-europe.com



Appendix DEMKO CERTIFICATE

Certificate No. D-06890-M1
Page 2/3
Date of Issue 2019-05-10

Model Details:

SE240AXXYZWW where S can be M or T, XX represents the output voltage from 12V to 48V, Y represents any number from 0 to 9 or blank, Z represents F, N, Q, while F means class I construction and N or Q means Class II construction, W represents any number from 0 to 9 or blank

Production Sites:

SL XIANGHE POWER ELECTRONICS CORP
No.B-02-03,North side of Landscape Ave, Qibu District,Environmental Industrial Park Xianghe, Hebei 065400
CHINA

Ratings:

Input: 100-240V~, 50-60Hz, 2.8A
Output: See Model Differences for the output ratings

Model ME240A1251Z01:
Input: 100-110Vac, 50-60Hz, 2.8A; Output: 12Vdc, 16.6A
Input: 110-240Vac, 50-60Hz, 2.8A; Output: 12Vdc, 16.6A

Model ME240A2451Z01:
Input: 100-110Vac, 50-60Hz, 2.8A; Output: 24Vdc, 9.17A
Input: 110-240Vac, 50-60Hz, 2.8A; Output: 24Vdc, 10.0A

Model ME240A2851Z01:
Input: 100-110Vac, 50-60Hz, 2.8A; Output: 28Vdc, 7.86A
Input: 110-240Vac, 50-60Hz, 2.8A; Output: 28Vdc, 8.60A

Model ME240A4851Z01:
Input: 100-110Vac, 50-60Hz, 2.8A; Output: 48Vdc, 5.0A
Input: 110-240Vac, 50-60Hz, 2.8A; Output: 48Vdc, 5.0A

Model ME240A2451F02:
Input: 100-110Vac, 50-60Hz, 2.8A; Output: 25Vdc, 8.80A
Input: 110-240Vac, 50-60Hz, 2.8A; Output: 25Vdc, 9.60A

Model ME240A1251Q02:
Input: 100-240Vac, 50-60Hz, 2.8A; Output: 12Vdc, 16.6A

Model ME240B2400F02:
Input: 100-110Vac, 50-60Hz, 2.8A; Output: 24Vdc, 9.17A
Input: 110-240Vac, 50-60Hz, 2.8A; Output: 24Vdc, 10.0A

Additional Information:

This certificate replaces certificate No. D-06890-UL dated 2018-11-08 due to issuing Technical Amendment to add a new model and component.


Certification Body

UL International Demko A/S, Borupvang 5A, DK-2750
Ballerup, Denmark, Tel. +45 44 85 65 65, info.dk@ul.com
www.ul-europe.com



Appendix DEMKO CERTIFICATE

Certificate No. D-06890-M1
Page 3/3
Date of Issue 2019-05-10



Certification Body

UL International Demko A/S, Borupvang 5A, DK-2750
Ballerup, Denmark, Tel. +45 44 85 65 65, info.dk@ul.com
www.ul-europe.com

