

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

CB TEST CERTIFICATE

Product

Switched Mode Power Supply

Name and address of the applicant

Astec International Ltd
16TH FLLU PLAZA2 WING YIP STKWUN TONG KOWLOON
Hongkong

Name and address of the manufacturer

EXCELSYS TECHNOLOGIES LTD
UNIT 27EASTGATE BUSINESS PARK LITTLE ISLAND
CORK IRELAND

Name and address of the factory

SHENZHEN WATT ELECTRONICS CO. LTD
NO. 5 TUNNEL 1, TANGFANG GARDEN, 35 DISTRICT, BAOAN
518101 SHENZHEN, GUANGDONG CHINA

Note: When more than one factory, please report on page 2

 Additional Information on page 2

Ratings and principal characteristics

Input: 100-240V, 50-60Hz, 5A-2.5A, Class I, no applied part.
Output: 600W maximum (see Table 2, Table 3, Table 4 in General
Information in Test Report).

Trademark / Brand (if any)



Type of Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

CX06M-wxyz-defgh (CoolX CoolPac with CoolMod);
CX06M-0000-defgh (CoolX CoolPac without CoolMod);
CX06M-0DC0-X-A-A (complete configuration)
The report was revised to include technical modifications.

Additional information (if necessary may also be reported on page 2)

 Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005

As shown in the Test Report Ref. No. which forms part of this Certificate

E182560-D1006-1/A1/C0-CB issued on 2020-12-28

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2021-01-07

Signature:

Original Issue Date: 2020-11-16

Jan-Erik Storgaard

Additional Information:

The original report was modified to include following changes/additions:
- update of LoCC.

National Differences specified in the CB Test Report.

The product has been evaluated with the exception of Risk Management assessment.

Additional information (if necessary)

UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA

UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK

UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN

UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/nbnames

Date: 2021-01-07

Original Issue Date: 2020-11-16

Signature:

Jan-Erik Storgaard